

REFERENCES

1. Fan JL; Patibandla SA; Kimura S; Rao TN; Desai RK; Seetharamaiah GS; Kurosky A; Prabhakar. BS Purification and characterization of a recombinant human thyroid peroxidase expressed in insect cells. J Autoimmun 1996; 9(4):529-36.
2. Feldt-Rasmussen U. Analytical and clinical performance goals for testing autoantibodies to thyroperoxidase, thyroglobulin, and thyrotropin receptor. Clin Chem 1996;42:160-3.
3. Franke WG; Schimming C; Wunderlich G. Can thyroid peroxidase be used as a complementary tumor marker besides thyroglobulin? Preliminary experience with determination of TPO in differentiated thyroid carcinomas. Anticancer Res 1997; 17(4B):2999-3002.
4. Haapala AM; Hyöty H; Parkkonen P; Mustonen J; Soppi E. Antibody reactivity against thyroid peroxidase and myeloperoxidase in autoimmune thyroiditis and systemic vasculitis. Scand J Immunol 1997; 46(1):78-85.
5. Mariotti S, Caturegli P, Piccolo P, Barbesino G, Pinchera A. Anti-thyroid peroxidase autoantibodies in thyroid diseases. J Clin Endocrinol Metab 1990;71:661-9.
6. Massart C, Guilhem I, Gibassier J, Allannire H, Nicol M. Comparison of thyroperoxidase and microsomal antibody assays in sera from patients with Graves disease. Clin Chem 1991;37:1777-80.
7. Nakamura H; Genma R; Mikami T; Kitahara A; Natsume H; Andoh S; Nagasawa S; Nishiyama K; Chida K; Sato A; Yoshimi T. High incidence of positive autoantibodies against thyroid peroxidase and thyroglobulin in patients with sarcoidosis. Clin Endocrinol (Oxf) 1997; 46(4):467-72.
8. Roti E, Gardini E, Minelli R, Bianconi L, Braverman LE. Prevalence of anti-thyroid peroxidase antibodies in serum in the elderly: comparison with other tests for anti-thyroid antibodies. Clin Chem 1992;38:88-92.

2025-08-08



Thyroid Peroxidase IgG ELISA

Catalog No. TP069G (96 Tests)

INTENDED USE

The Calbiotech Thyroid Peroxidase (TPO) IgG ELISA Kit is intended for the detection of IgG antibody to Thyroid Peroxidase in human serum or plasma. **For Research Use Only. For professional use only. Not for use in diagnostic procedures.**

SUMMARY AND EXPLANATION

Thyroid peroxidase (TPO) is an enzyme critical to thyroid hormone synthesis. IgG antibodies to TPO are of research interest in studies on autoimmune thyroid disease, immune tolerance, and genetic susceptibility to autoimmunity.

PRINCIPLE OF THE TEST

Diluted patient serum is added to wells coated with purified TPO recombinant antigen. TPO IgG specific antibody, if present, binds to the antigen. All unbound materials are washed away, and the enzyme conjugate is added to bind to the antibody-antigen complex, if present. Excess enzyme conjugate is washed off and substrate is added. The plate is incubated to allow the hydrolysis of the substrate by the enzyme. The intensity of the color generated is proportional to the amount of IgG specific antibody in the sample.

MATERIALS PROVIDED		96 TESTS
1.	Microwells coated with TPO recombinant Ag	12x8x1
2.	Sample Diluent: 1 bottle (ready to use)	22 mL
3.	Calibrator: 1 Vial (ready to use)	1 mL
4.	Positive Control: 1 vial (ready to use)	1 mL
5.	Negative Control: 1 vial (ready to use)	1 mL
6.	Enzyme conjugate: 1 bottle (ready to use)	12 mL
7.	TMB Substrate: 1 bottle (ready to use)	12 mL
8.	Stop Solution: 1 bottle (ready to use)	12 mL
9.	Wash Concentrate 20X: 1 bottle	25 mL

MATERIALS NOT PROVIDED

1. Distilled or deionized water
2. Precision pipettes. Disposable pipette tips
3. ELISA reader capable of reading absorbance at 450nm
4. Absorbance paper or paper towel

STORAGE AND STABILITY

1. Store the kit at 2-8°C.
2. Keep microwells sealed in a dry bag with desiccants.
3. The reagents are stable until expiration of the kit.
4. Do not expose test reagents to heat, sun or strong light during storage or usage.

**Calbiotech, Inc.**

1935 Cordell Ct., El Cajon, CA 92020 USA

Tel (619) 660-6162 | Fax (619) 660-6970 | Web www.calbiotech.com

WARNINGS AND PRECAUTIONS

Potential biohazardous materials:

1. The standards contain human source components which have been tested and found non-reactive for hepatitis B surface antigen as well as HIV antibody with FDA licensed reagents. However, as there is no test method that can offer complete assurance that HIV, Hepatitis B virus or other infectious agents are absent, these reagents should be handled at the Biosafety Level 2, as recommended in the Centers for Disease Control/National Institutes of Health manual, "Biosafety in Microbiological and Biomedical Laboratories." 1984.
2. Do not pipette by mouth. Do not smoke, eat, or drink in the areas in which specimens or kit reagents are handled.
3. The components in this kit are intended for use as an integral unit. The components of different lots should not be mixed.
4. It is recommended that standards, control and serum samples be run in duplicate
5. Optimal results will be obtained by strict adherence to this protocol. Accurate and precise pipetting, as well as following the exact time and temperature requirements prescribed are essential. Any deviation from this may yield invalid data.

SPECIMEN COLLECTION AND HANDLING

This assay is designed for use with human serum or plasma samples obtained in compliance with applicable laws, regulations, and institutional policies. Handle and store samples using procedures appropriate for research use. Samples may be stored refrigerated (2–8 °C) for up to seven days, or frozen (–20 °C or below) for up to six months. Avoid repetitive freeze–thaw cycles.

REAGENT PREPARATION

Prepare 1X Wash buffer by adding the contents of the bottle (25 ml, 20X) to 475 ml of distilled or deionized water. Store at room temperature (20-25°C).

ASSAY PROCEDURE

Bring all specimens and kit reagents to room temperature (20-25°C) and gently mix.

1. Place the desired number of coated strips into the holder.
2. Negative control, positive control, and calibrator are ready to use. Prepare 1:21 dilution of test samples, by adding 10 µl of the sample to 200 µl of sample diluent. Mix well.
3. Dispense 100 µl of diluted sera, calibrator and controls into the appropriate wells. For the reagent blank, dispense 100µl sample diluent in 1A well position. Tap the holder to remove air bubbles from the liquid and mix well. Incubate for 20 minutes at room temperature.
4. Remove liquid from all wells. Wash wells three times with 300 µl of 1X wash buffer. Blot on absorbance paper or paper towel.
5. Dispense 100 µl of enzyme conjugate to each well and incubate for 20 minutes at room temperature.
6. Remove enzyme conjugate from all wells. Wash wells three times with 300 µl of 1X wash buffer. Blot on absorbance paper or paper towel.
7. Dispense 100 µl of TMB substrate and incubate for 10 minutes at room temperature.
8. Add 100 µl of stop solution.
9. Read O.D. at 450 nm using ELISA reader within 15 min. A dual wavelength is recommended with reference filter of 600-650 nm.

CALCULATION OF RESULTS

1. Check Calibrator Factor (CF) value on the calibrator bottle. This value might vary from lot to lot. Make sure you check the value on every kit.
2. Calculate cut-off value: Calibrator OD x Calibrator Factor (CF).
3. Calculate the Ab (Antibody) Index of each determination by dividing the mean values of each sample by cut-off value.

QUALITY CONTROL

The test run may be considered valid provided the following criteria are met:

1. The O.D. of the Calibrator should be greater than 0.250.
2. The Ab index for Negative control should be less than 0.9.
3. The Ab Index for Positive control should fall within the range specified on the COA/label.

LIMITATIONS OF THE TEST

1. Lipemic or hemolyzed samples may cause erroneous results.